

Unit: micron

Peak %T = 90.2% at 4.342

Half power points: 45.1%T @ 4.278 45.1%T @ 4.434

Half band width = 0.155 % Band width = 3.574%


Center wavelength = 4.356

Test conditions: temp=amb. angle=0 PE9830

Instrument operating parameters: res=3.0cm-1 apert=.8"

Attenuator=0.1%T Below-BP-8.0um

Date: 05/22/91



FILTER SPECIFICATION

- 112 Narrow Band CO/2 Analysis



© ELTEC Instruments, Inc.

Download from the ELTEC Information Center at <http://www.silverlight.ch>

Updated Jan. 2002